



Nederlands Meetinstituut

OIML Member State
The Netherlands

OIML Certificate N° R76/1992-NL1-03.62

Project number 213685

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: Ing. C. Oosterman

Applicant

Name: A&D Instruments Ltd.
Address: 24 Blacklands Way
Abingdon Business Park
Abingdon, Oxfordshire
OX14 1DY United Kingdom

Manufacturer of the certified type

Name: A&D Company Ltd.
Address: 3-23-14 Higashi-Ikebukuro
Toshima-Ku
Tokyo 170 Japan

Identification of the certified type

Type : EW-...i

 $30 \text{ g} \leq \text{Max} \leq 12 \text{ kg}$
 $e \geq 0.02 \text{ g}$
 $n \leq 3000$ divisions (per weighing range)
maximum of 3 weighing ranges

Temperature range +5 °C / +40 °C

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6391 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
Edition 1992

for accuracy class (II) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of the instrument covered by the relevant OIML International Recommendation identified above.

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V.
(Chamber of Commerce no.27.228.701)

Subsidiary companies:
NMI Van Swinden Laboratorium B.V. (27228703)
NMI Certin B.V. (27.233.418)
Verispect B.V. (27.228.700)

This document is issued under the provision that NMI. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



Nederlands Meetinstituut

OIML Member State
The Netherlands

OIML Certificate N° R76/1992-NL1-03.62

Project number 213685
Page 2 of 2

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

- N° R76/1992-NL1-03.62C2 that includes 8 pages;
- N° R76/1992-NL1-03.62C2W1 that includes 22 pages;
- N° R76/1992-NL1-03.62C2W2 that includes 22 pages;
- N° R76/1992-NL1-03.62C2W3 that includes 47 pages;
- N° R76/1992-NL1-03.62C3 that includes 16 pages;
- N° R76/1992-NL1-03.62C3W1 that includes 8 pages;
- N° R76/1992-NL1-03.62C3W2 that includes 8 pages;
- N° R76/1992-NL1-03.62C3W3 that includes 8 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

17 December 2003

The CIML Member
G.J. Faber

17 December 2003

*
**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.